

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1 mm	1.001 mm	1 µm	± 4 µm	±	5.8E-01 µm
1.25 mm	1.251 mm	1 µm	± 4 µm	±	5.8E-01 µm
1.5 mm	1.500 mm	1 µm	± 4 µm	±	5.8E-01 µm
2 mm	2.001 mm	1 µm	± 4 µm	±	5.8E-01 µm
3 mm	3.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
4 mm	4.000 mm	0 µm	± 4 µm	±	5.9E-01 µm
5 mm	5.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
10 mm	10.000 mm	0 µm	± 4 µm	±	5.8E-01 µm
15 mm	15.001 mm	1 µm	± 4 µm	±	5.8E-01 µm
20 mm	20.000 mm	0 µm	± 4 µm	±	5.9E-01 µm
25 mm	25.000 mm	0 µm	± 4 µm	±	5.9E-01 µm