

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.60 μm	51.16 μm	0.56 μm	± 2.5 μm	±	5.8E-02 μm
120.30 μm	120.2 μm	-0.10 μm	± 2.5 μm	±	6.1E-02 μm
246.20 μm	246.4 μm	0.20 μm	± 3.8 μm	±	6.9E-02 μm
479.00 μm	479.0 μm	0.00 μm	± 14.3 μm	±	9.3E-02 μm