

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.180 °C	0.0 °C	-0.180 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.8 °C	-0.235 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.377 °C	35.1 °C	-0.277 °C	± 0.5 °C	± 4.0E-01 °C