

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.135 °C	1.8 °C	-0.335 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	4.211 °C	4.1 °C	-0.111 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	6.116 °C	5.7 °C	-0.416 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.052 °C	7.9 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C