

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0244 µl	10 µl	-0.024 µl	± 0.1 µl	±	5.1E-02 µl
50.0219 µl	50 µl	-0.022 µl	± 0.1 µl	±	5.2E-02 µl
100.044 µl	100 µl	-0.044 µl	± 0.1 µl	±	5.3E-02 µl