

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.051 °C	15.3 °C	0.249 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.2 °C	0.172 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.5 °C	0.287 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.018 %HR	64.7 %HR	4.682 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	70.0 %HR	4.665 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	74.3 %HR	3.852 %HR	± 3.0 %HR	± 1.9E+00 %HR