

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.051 °C	15.2 °C	0.149 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.2 °C	0.172 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.3 °C	0.087 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.018 %HR	65.0 %HR	4.982 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.335 %HR	69.8 %HR	4.465 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	75.2 %HR	4.752 %HR	± 3.0 %HR	± 1.9E+00 %HR