

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.051 °C                  | 15.3 °C            | 0.249 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.028 °C                  | 25.3 °C            | 0.272 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.213 °C                  | 30.3 °C            | 0.087 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 60 %HR                     | 60.018 %HR                 | 64.8 %HR           | 4.782 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 65 %HR                     | 65.335 %HR                 | 69.8 %HR           | 4.465 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.448 %HR                 | 74.4 %HR           | 3.952 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |