

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -30 °C            | -30.504 °C        | -30.1 °C  | 0.404 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| -25 °C            | -25.290 °C        | -24.9 °C  | 0.390 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| -15 °C            | -15.363 °C        | -15.2 °C  | 0.163 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| -10 °C            | -10.148 °C        | -9.7 °C   | 0.448 °C      | ± 3.0 °C             | ± 4.0E-01 °C                |
| 2 °C              | 1.985 °C          | 2.1 °C    | 0.115 °C      | ± 3.0 °C             | ± 4.0E-01 °C                |
| 8 °C              | 8.102 °C          | 8.0 °C    | -0.102 °C     | ± 3.0 °C             | ± 4.0E-01 °C                |