

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-30 °C	-30.504 °C	-30.1 °C	0.404 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.290 °C	-24.9 °C	0.390 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.363 °C	-15.2 °C	0.163 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.148 °C	-9.7 °C	0.448 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	2.1 °C	0.115 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.102 °C	8.0 °C	-0.102 °C	± 0.5 °C	± 4.0E-01 °C