

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-30 °C	-30.504 °C	-30.1 °C	0.404 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.290 °C	-25.0 °C	0.290 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.363 °C	-15.1 °C	0.263 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.148 °C	-9.9 °C	0.248 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.985 °C	1.8 °C	-0.185 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.102 °C	8.0 °C	-0.102 °C	± 0.5 °C	± 4.0E-01 °C