

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
751.00 μm	751.80 μm	0.80 μm	± 2.5 μm	±	1.3E-01 μm
1,494.00 μm	1,495.5 μm	1.50 μm	± 2.5 μm	±	2.4E-01 μm