

Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
7 Hz	0.551 m/s^2-rms	0.64 m/s^2-rms	0.09 m/s^2-rms		±	0.087 m/s^2-rms
7 Hz	1.523 m/s^2-rms	1.64 m/s^2-rms	0.11 m/s^2-rms		±	0.087 m/s^2-rms
7 Hz	2.771 m/s^2-rms	2.95 m/s^2-rms	0.18 m/s^2-rms		±	0.087 m/s^2-rms
10 Hz	3.043 m/s^2-rms	3.16 m/s^2-rms	0.12 m/s^2-rms		±	0.127 m/s^2-rms
10 Hz	5.543 m/s^2-rms	5.81 m/s^2-rms	0.27 m/s^2-rms		±	0.127 m/s^2-rms
30 Hz	6.936 m/s^2-rms	7.31 m/s^2-rms	0.37 m/s^2-rms		±	0.155 m/s^2-rms
50 Hz	6.936 m/s^2-rms	7.33 m/s^2-rms	0.39 m/s^2-rms		±	0.152 m/s^2-rms
80 Hz	6.936 m/s^2-rms	7.41 m/s^2-rms	0.47 m/s^2-rms		±	0.153 m/s^2-rms
100 Hz	6.936 m/s^2-rms	7.32 m/s^2-rms	0.38 m/s^2-rms		±	0.154 m/s^2-rms