

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
30 Hz	1.000 g-pk	1.005 g-pk	0.005 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
50 Hz	1.000 g-pk	1.002 g-pk	0.002 g-pk	± 0.01 g-pk	±	2.0E-02 g-pk
80 Hz	1.000 g-pk	1.004 g-pk	0.004 g-pk	± 0.01 g-pk	±	2.0E-02 g-pk
100 Hz	1.000 g-pk	0.999 g-pk	-0.001 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk
500 Hz	1.000 g-pk	1.002 g-pk	0.002 g-pk	± 0.01 g-pk	±	2.0E-02 g-pk
1000 Hz	1.000 g-pk	1.001 g-pk	0.001 g-pk	± 0.01 g-pk	±	2.1E-02 g-pk