

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-20 °C	-20.798 °C	-20.5 °C	0.298 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.020 °C	0.5 °C	0.480 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	7.881 °C	8.0 °C	0.119 °C	± 1.0 °C	± 4.0E-01 °C