

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.481 °C	29.4 °C	-0.081 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.079 %HR	28.7 %HR	-1.379 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.828 %HR	50.5 %HR	-1.328 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.139 %HR	73.8 %HR	-0.339 %HR	± 3.0 %HR	± 1.9E+00 %HR