

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.092 °C	30.2 °C	0.108 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	38.0 %HR	2.829 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	49.3 %HR	-0.933 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.3 %HR	-1.943 %HR	± 3.0 %HR	± 1.9E+00 %HR