

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 1.7E-01 °C
36 °C	36.179 °C	36.4 °C	0.221 °C	± 0.5 °C	± 1.7E-01 °C
37 °C	37.383 °C	37.6 °C	0.217 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	40.090 °C	40.4 °C	0.310 °C	± 0.5 °C	± 1.6E-01 °C