

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
23 °C	23.043 °C	23.3 °C	0.257 °C	± 0.5 °C	± 1.7E-01 °C
25 °C	25.450 °C	25.2 °C	-0.250 °C	± 0.5 °C	± 1.7E-01 °C
30 °C	30.062 °C	30.4 °C	0.338 °C	± 0.5 °C	± 1.6E-01 °C
36 °C	36.179 °C	36.4 °C	0.221 °C	± 0.5 °C	± 1.6E-01 °C
37 °C	37.383 °C	37.5 °C	0.117 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	40.090 °C	40.4 °C	0.310 °C	± 0.5 °C	± 1.6E-01 °C