

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	30.5 °C	0.207 °C	± 0.5 °C	± 1.6E-01 °C
24 %HR	24.435 %HR	26.0 %HR	1.565 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	60.223 %HR	57.9 %HR	-2.323 %HR	± 5.0 %HR	± 1.9E+00 %HR
90 %HR	90.182 %HR	91.0 %HR	0.818 %HR	± 5.0 %HR	± 1.9E+00 %HR