

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0244 µl	10 µl	-0.024 µl	± 0.12 µl	±	5.1E-02 µl
49.9217 µl	50 µl	0.078 µl	± 0.5 µl	±	5.2E-02 µl
99.843 µl	100 µl	0.157 µl	± 0.8 µl	±	5.3E-02 µl