

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.8433 µl	100 µl	0.157 µl	± 0.8 µl	±	5.3E-02 µl
501.3216 µl	500 µl	-1.322 µl	± 4 µl	±	5.9E-02 µl
1001.039 µl	1000 µl	-1.039 µl	± 8 µl	±	7.8E-02 µl