

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.4952 ml	0.5 ml	0.005 ml	± 4 ml	±	2.0E+01 µl
2.4921 ml	2.5 ml	0.007927 ml	± 40 ml	±	2.0E+01 µl
4.984146 ml	5 ml	0.015854 ml	± 40 ml	±	2.0E+01 µl