

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.4952 ml	0.5 ml	0.0048 ml	± 0.01 ml	±	5.9E-02 µl
2.4921 ml	2.5 ml	0.0079 ml	± 0.03 ml	±	1.6E-01 µl
4.984146 ml	5 ml	0.0159 ml	± 0.03 ml	±	3.0E-01 µl