

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0932 ml	0.1 ml	0.0068 ml	± 0.01 ml	±	5.1E-02 µl
0.5016 ml	0.5 ml	-0.0016 ml	± 0.01 ml	±	5.4E-02 µl
1.003691 ml	1 ml	-0.0037 ml	± 0.01 ml	±	8.1E-02 µl