

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.942 °C	30.1 °C	0.158 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	34.5 %HR	2.488 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.818 %HR	54.8 %HR	2.982 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.306 %HR	75.1 %HR	-1.206 %HR	± 3.0 %HR	± 1.9E+00 %HR