

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.942 °C	30.3 °C	0.358 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	34.5 %HR	2.488 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.818 %HR	54.7 %HR	2.882 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.306 %HR	75.3 %HR	-1.006 %HR	± 3.0 %HR	± 1.9E+00 %HR