

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.063 in	0.0000 in	± 0.0002 in	±	7.3E-05 in
0.100 in	0.100 in	0.0000 in	± 0.0002 in	±	9.1E-05 in
0.125 in	0.125 in	0.0000 in	± 0.0002 in	±	1.0E-04 in
0.200 in	0.200 in	0.0000 in	± 0.0002 in	±	1.5E-04 in
0.250 in	0.250 in	0.0001 in	± 0.0002 in	±	1.8E-04 in
0.300 in	0.300 in	0.0000 in	± 0.0002 in	±	2.2E-04 in
0.313 in	0.313 in	0.0000 in	± 0.0002 in	±	2.3E-04 in
0.350 in	0.350 in	-0.0001 in	± 0.0002 in	±	2.5E-04 in
0.400 in	0.400 in	0.0000 in	± 0.0002 in	±	2.9E-04 in