

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.253 °C	30.1 °C	-0.153 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.478 %HR	29.4 %HR	-1.078 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.333 %HR	46.9 %HR	-2.433 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.814 %HR	65.3 %HR	-4.514 %HR	± 5.0 %HR	± 1.9E+00 %HR