

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.253 °C	29.6 °C	-0.653 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.478 %HR	27 %HR	-3.478 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.333 %HR	47 %HR	-2.333 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.814 %HR	67 %HR	-2.814 %HR	± 5.0 %HR	± 1.9E+00 %HR