

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.253 °C	29.7 °C	-0.553 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.478 %HR	24 %HR	-6.978 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.333 %HR	43 %HR	-6.433 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.814 %HR	64 %HR	-5.414 %HR	± 5.0 %HR	± 1.9E+00 %HR