

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
°C	30.112 °C	30.0 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	27.9 %HR	-2.670 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.386 %HR	52.2 %HR	1.814 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	69.5 %HR	-0.539 %HR	± 3.0 %HR	± 1.9E+00 %HR