

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.942 °C	29.8 °C	-0.142 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	27 %HR	-5.012 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.818 %HR	47 %HR	-4.818 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	76.306 %HR	73 %HR	-3.306 %HR	± 5.0 %HR	± 1.9E+00 %HR