

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99950	0.500	± 1	±	4.3E-01 mm						
0 a 2	1.99952	0.480	± 1.4	±	4.8E-01 mm						
0 a 3	2.99927	0.730	± 1.8	±	4.8E-01 mm						
0 a 4	3.99965	0.350	± 2.2	±	4.8E-01 mm						
0 a 5	4.99898	1.020	± 2.6	±	4.8E-01 mm						
0 a 6	5.99928	0.720	± 3	±	3.6E-01 mm						
0 a 7	6.99963	0.370	± 3.4	±	3.6E-01 mm						
0 a 8	7.99920	0.800	± 3.8	±	3.6E-01 mm						
0 a 9	8.99938	0.620	± 4.2	±	3.6E-01 mm						
0 a 10	9.99983	0.170	± 4.6	±	3.6E-01 mm						