

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60.071 s	60.04 s	-0.031 s	-0.0516%	± 0.1%	±	5.0E-02 s
60.317 s	60.28 s	-0.037 s	-0.0613%	± 0.1%	±	5.0E-02 s
60.253 s	60.21 s	-0.043 s	-0.0714%	± 0.1%	±	5.0E-02 s