

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
12.4 RPM	12 RPM	-0.4 RPM	± 0.01 RPM	±	9.4E-02 RPM
29.9 RPM	30 RPM	0.1 RPM	± 0.03 RPM	±	9.4E-02 RPM
59.9 RPM	60 RPM	0.1 RPM	± 0.06 RPM	±	9.4E-02 RPM
99.9 RPM	100 RPM	0.1 RPM	± 0.10 RPM	±	5.8E+01 RPM