

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.000 mm	0.000 μm	± 4 μm	±	5.8E-01 μm
1.25 mm	1.250 mm	0.000 μm	± 4 μm	±	5.8E-01 μm
1.50 mm	1.500 mm	0.000 μm	± 4 μm	±	5.8E-01 μm
2.00 mm	2.000 mm	0.000 μm	± 4 μm	±	5.8E-01 μm
3.00 mm	3.000 mm	0.000 μm	± 4 μm	±	5.8E-01 μm
5.00 mm	5.001 mm	1.000 μm	± 4 μm	±	5.8E-01 μm
10.00 mm	10.001 mm	1.000 μm	± 4 μm	±	5.8E-01 μm
15.00 mm	15.001 mm	1.000 μm	± 4 μm	±	5.8E-01 μm
20.00 mm	20.001 mm	1.000 μm	± 4 μm	±	5.8E-01 μm
25.00 mm	25.001 mm	1.000 μm	± 4 μm	±	5.8E-01 μm