

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.211 °C	15.2 °C	-0.011 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.932 °C	30.1 °C	0.168 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	32.1 %HR	0.088 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.153 %HR	68.4 %HR	-2.753 %HR	± 3.0 %HR	± 1.9E+00 %HR