

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.211 °C	15.1 °C	-0.111 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.932 °C	30.1 °C	0.168 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	32.012 %HR	32.2 %HR	0.188 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.153 %HR	69.1 %HR	-2.053 %HR	± 3.0 %HR	± 1.9E+00 %HR