

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.298 °C	11.5 °C	1.202 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.416 °C	20.5 °C	0.084 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.253 °C	29.5 °C	-0.753 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.010 °C	39.2 °C	-0.810 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	25.355 %HR	22 %HR	-3.355 %HR	± 5.0 %HR	± 1.9E+00 %HR
40 %HR	40.079 %HR	39 %HR	-1.079 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.619 %HR	64 %HR	4.381 %HR	± 5.0 %HR	± 1.9E+00 %HR