

Certificado No. (Certificate #):62467Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-21Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
2.50 mm	2.50 mm	0 µm	± 4 µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
7.00 mm	7.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
12.00 mm	12.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
15.00 mm	15.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
17.00 mm	17.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
20.00 mm	20.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
23.00 mm	23.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
25.00 mm	25.00 mm	0 µm	± 4 µm	±	5.8E+00 µm