

Certificado No. (Certificate #):62468Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-21Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
25.00 mm	25.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
27.50 mm	27.50 mm	0 µm	± 4 µm	±	5.8E+00 µm
30.00 mm	30.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
32.00 mm	32.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
35.00 mm	35.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
37.00 mm	37.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
40.00 mm	40.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
43.00 mm	43.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
45.00 mm	45.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
48.00 mm	48.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
50.00 mm	50.00 mm	0 µm	± 4 µm	±	5.8E+00 µm