

Certificado No. (Certificate #):62469Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-21Marca (Brand):INSIZE

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 50.00 mm          | 50.00 mm  | 0 μm          | ± 4 μm            | ±                           | 5.8E+00 μm |
| 53.00 mm          | 53.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 55.00 mm          | 55.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 58.00 mm          | 58.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 60.00 mm          | 60.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 62.00 mm          | 62.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 65.00 mm          | 65.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 67.00 mm          | 67.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 70.00 mm          | 70.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 72.00 mm          | 72.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |
| 75.00 mm          | 75.00 mm  | 0 μm          | ± 5 μm            | ±                           | 5.8E+00 μm |