

Certificado No. (Certificate #):62470Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-21Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
75.00 mm	75.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
77.00 mm	77.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
80.00 mm	80.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
82.00 mm	82.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
85.00 mm	85.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
88.00 mm	88.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
90.00 mm	90.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
93.00 mm	93.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
95.00 mm	95.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
97.00 mm	97.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
100.00 mm	100.00 mm	0 μm	± 5 μm	±	5.8E+00 μm