

Certificado No. (Certificate #):62471Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-22Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
100.00 mm	100.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
102.00 mm	102.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
105.00 mm	105.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
108.00 mm	108.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
110.00 mm	110.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
113.00 mm	113.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
115.00 mm	115.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
117.00 mm	117.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
120.00 mm	120.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
123.00 mm	123.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
125.00 mm	125.00 mm	0 µm	± 6 µm	±	5.8E+00 µm