

Certificado No. (Certificate #):62471Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-22Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
125.00 mm	125.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
127.00 mm	127.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
130.00 mm	130.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
132.00 mm	132.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
135.00 mm	135.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
137.00 mm	137.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
140.00 mm	140.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
143.00 mm	143.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
145.00 mm	145.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
148.00 mm	148.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
150.00 mm	150.00 mm	0 μm	± 6 μm	±	5.8E+00 μm