

Certificado No. (Certificate #):62473Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-22Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
150.00 mm	150.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
152.00 mm	152.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
155.00 mm	155.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
158.00 mm	158.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
160.00 mm	160.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
163.00 mm	163.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
165.00 mm	165.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
167.00 mm	167.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
170.00 mm	170.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
173.00 mm	173.00 mm	0 μm	± 7 μm	±	5.8E+00 μm
175.00 mm	175.00 mm	0 μm	± 7 μm	±	5.8E+00 μm