

Certificado No. (Certificate #):62474Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-22Marca (Brand):INSIZE

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 175.00 mm         | 175.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 177.00 mm         | 177.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 180.00 mm         | 180.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 183.00 mm         | 183.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 185.00 mm         | 185.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 188.00 mm         | 188.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 190.00 mm         | 190.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 192.00 mm         | 192.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 195.00 mm         | 195.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 198.00 mm         | 198.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |
| 200.00 mm         | 200.00 mm | 0 µm          | ± 7 µm            | ±                           | 5.8E+00 µm |