

Certificado No. (Certificate #):62475Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-23Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
200.00 mm	200.00 mm	0 μm	± 7 μm	±	5.9E+00 μm
202.00 mm	202.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
205.00 mm	205.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
208.00 mm	208.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
210.00 mm	210.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
213.00 mm	213.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
215.00 mm	215.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
217.00 mm	217.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
220.00 mm	220.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
223.00 mm	223.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
225.00 mm	225.00 mm	0 μm	± 8 μm	±	5.9E+00 μm