

Certificado No. (Certificate #):62476Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-23Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
225.00 mm	225.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
228.00 mm	228.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
230.00 mm	230.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
233.00 mm	233.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
235.00 mm	235.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
237.00 mm	237.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
240.00 mm	240.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
242.00 mm	242.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
245.00 mm	245.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
248.00 mm	248.00 mm	0 μm	± 8 μm	±	5.9E+00 μm
250.00 mm	250.00 mm	0 μm	± 8 μm	±	5.9E+00 μm